



# CERTIFIKAT

No. S-28 46 02 rev 1

## Non-automatic Weighing Instrument – S500 and SK500

### Holder/Issued to

Maxicap AB, Tallskogsvägen 11, 793 35 Leksand, Sweden

### Product name

S500 and SK500

### Description

A graduated, self-indicating, electronic, single-interval non-automatic weighing instrument, intended to not be used for direct sales to the public

Manufacturer	Maxicap AB	
Data	Accuracy class	III
	Maximum capacity	≤ 500 kg
	Verification scale interval	e ≥ 0,2 kg
	Number of verification scale intervals	500 ≤ n ≤ 2500

The principal characteristics and approval conditions are set out in the appendix to this certificate, which forms part of the approval document.

### Validity


This certificate is valid until not later than 14<sup>th</sup> March 2017. This certificate was first issued 15<sup>th</sup> March 2007, and this is revision 1 of the certificate.


### Certificate issued by

SP Technical Research Institute of Sweden, Box 857, SE-501 15 Borås, Sweden, Notified Body Number 0402, in accordance with the regulations STAFS 2007:18 on non-automatic weighing instruments which implement, in Sweden, the Council Directive 90/384/EEC.

Borås, 15<sup>th</sup> December 2010

### SP Technical Research Institute of Sweden Certification

  
Lennart Aronsson  
Product Certification Manager

  
Bengt Gutfelt  
Certification Officer

Appendix (7 pages)

#### SP Technical Research Institute of Sweden

Postal address Phone / Fax Reg.number E-mail / Internet  
Box 857 +46 10 516 50 00 556464-6874 info@sp.se  
SE-501 15 Borås +46 33 13 55 02 www.sp.se  
SWEDEN

#### EC Type Approval Certificate issued by Notified Body No. 0402

Swedish Notified Bodies are appointed by SWEDAC, the Swedish Board for Accreditation and Conformity Assessment, under the terms of Swedish legislation. This certificate may not be reproduced other than in full, except with the prior written approval by SP.