



CERTIFIKAT

No. S-28 46 01 rev 1

Non-automatic Weighing Instrument - PT 100

Holder/Issued to

Maxicap AB, Tallskogsvägen 11, 793 35 Leksand, Sweden

Product name

PT 100

Description

A graduated, self-indicating, electronic, multi-interval, non-automatic weighing instrument

Manufacturer Maxicap AB

Data	Accuracy class	III
	Maximum capacity	Max \leq 100 kg
	Verification scale interval	$e_i \geq$ 1 g
	Number of verification scale intervals	$500 \leq n_i \leq$ 3000

The principal characteristics, approval conditions and special conditions, if any, are set out in the Appendix hereto, which forms part of the approval document and consists of 3 pages.

Validity

This certificate is valid until not later than 31st December 2020. This certificate was first issued 20th December 2000 and this is revision 1 of the certificate.

Certificate issued by

SP Technical Research Institute of Sweden, Box 857, SE-501 15 Borås, Sweden, Notified Body Number 0402, in accordance with the regulations STAFS 2007:18 on non-automatic weighing instruments which implements, in Sweden, the Council Directive 90/384/EEC.

Borås, 15th December 2010

SP Technical Research Institute of Sweden Certification

Lennart Aronsson
Product Certification Manager

Bengt Gutfelt
Certification Officer

Appendix (4 pages)

SP Technical Research Institute of Sweden

Postal address Box 857 SE-501 15 Borås SWEDEN
Phone / Fax +46 10 516 50 00 +46 33 13 55 02
Reg. number 556464-6874
E-mail / Internet info@sp.se www.sp.se

EC Type Approval Certificate issued by Notified Body No. 0402

Swedish Notified Bodies are appointed by SWEDAC, the Swedish Board for Accreditation and Conformity Assessment, under the terms of Swedish legislation. This certificate may not be reproduced other than in full, except with the prior written approval by SP.